



## Certificate/Report

RESULTS REPORTED RELATED ONLY TO ITEMS TESTED

**COMPANY NAME** : FORTY THOUSAND AND 75KM (PTY) LTD  
**ADDRESS** : PRIVATE BAG X2007, MENLYN, PRETORIA, GAUTENG, 0063  
**SUBJECT** : QUANTITATIVE SUSPENSION TEST FOR THE EVALUATION OF  
BACTERICIDAL ACTIVITY OF CHEMICAL DISINFECTANTS  
AND ANTISEPTICS SANS 51276 (EN1276): DILUTION-NEUTRALIZATION  
METHOD  
**MARKED** : STERAMINE 6-Q TABLETS MULTI-PURPOSE SANITIZER 0113  
**ACTIVE INGREDIENT** : DIMETHYL BENZYL AMMONIUM CHLORIDE DIHYDRATE 200PPM  
**APPEARANCE** : LIGHT PURPLE TABLET  
**NEUTRALIZING AGENT** : NEUTRALIZING FLUID  
**STORAGE CONDITIONS** : STORE IN A COOL, DRY PLACE AWAY FROM DIRECT SUNLIGHT  
**INSTRUCTED BY** : VICTOR  
**LAB NO.** : P231331  
**RECEIVED ON** : 09/07/2020  
**DATE ANALYSED** : 09/07/2020

### EXPERIMENTAL CONDITIONS:

Obligatory conditions  
test organisms:

*Enterococcus hirae* ATCC 10541

*Escherichia coli* ATCC 10536

*Pseudomonas aeruginosa* ATCC 15442

*Staphylococcus aureus* ATCC 6538

Test temperature:

20°C

Contact time:

5 minutes

Interfering substance:

0.3g/l Bovine serum albumin (Clean conditions)

Test incubation temperature:

37°C

### PASS REQUIREMENTS:

The product shall demonstrate at least a 5 decimal log reduction when diluted with hard water/or undiluted and tested under obligatory test conditions. At least one of the test concentrations will demonstrate a log reduction of less than 5 log.

TECHNICAL SIGNATORY  
TINLABAZA

Authorised Signature (original blue ink)



## Certificate/Report

RESULTS REPORTED RELATED ONLY TO ITEMS TESTED

### TEST VALIDITY

For each test organism:

- $N$  is between  $1.5 \times 10^8$  and  $5.0 \times 10^8$  ( $8.17 \leq \log N \leq 8.70$ )
- $N_0$  is between  $1.5 \times 10^7$  and  $5.0 \times 10^7$  ( $7.17 \leq \log N_0 \leq 7.70$ )
- $N_{10}$  is between 30 and 160 ( $3.0 \times 10^1$  and  $1.6 \times 10^2$ )
- A, B, C are equal to or greater than  $0.5 \times N_{10}$
- Control of weighted mean counts: quotient is not lower than 5 and not higher than 15

### RESULTS:

#### Escherichia coli ATCC 10536

Table 1:  $N$  and  $N_0$  values

Dilution	Vc1	Vc2	Average N(wm)	Log N	$N_0$ (N/10)	Log $N_0$
$10^{-6}$	320	312	$3.8 \times 10^8$	8.58	$3.8 \times 10^7$	7.58
$10^{-7}$	42	48				
Is Log $N$ between 8.17 and 8.70: <b>Yes</b>						
Is Log $N_0$ between 7.17 and 7.70: <b>Yes</b>						
Control of weighted mean counts: <b>7.02</b>						

Table 2: Test Log Reduction Values

Product Concentration	Vc1	Vc2	$N_a$ (Ave Vc1&Vc2x10)	Log $N_a$	Log Reduction (No: 7.58)
1ppm	>330	>330	>3300	>3.52	<4.06
200ppm	<14	<14	<140	<2.15	>5.43

#### Pseudomonas aeruginosa ATCC 15442

Table 3:  $N$  and  $N_0$  values

Dilution	Vc1	Vc2	Average N(wm)	Log N	$N_0$ (N/10)	Log $N_0$
$10^{-6}$	304	294	$3.2 \times 10^8$	8.51	$3.2 \times 10^7$	7.51
$10^{-7}$	32	36				
Is Log $N$ between 8.17 and 8.70: <b>Yes</b>						
Is Log $N_0$ between 7.17 and 7.70: <b>Yes</b>						
Control of weighted mean counts: <b>8.79</b>						

TECHNICAL SIGNATORY  
 TINKABAZA

  
 .....  
 Authorised Signature (original blue ink)



M&L LABS

# Certificate/Report

RESULTS REPORTED RELATED ONLY TO ITEMS TESTED

**Table 4: Test Log Reduction Values**

Product Concentration	Vc1	Vc2	Na (Ave Vc1&Vc2x10)	Log Na	Log Reduction (No:7.51)
1ppm	>330	>330	>3300	>3.52	<3.99
200ppm	<14	<14	<140	<2.15	>5.36

**Staphylococcus aureus ATCC 6538**

**Table 5: N and N<sub>0</sub> values**

Dilution	Vc1	Vc2	Average N(wm)	Log N	N <sub>0</sub> (N/10)	Log N <sub>0</sub>
10 <sup>-6</sup>	320	333	3.7x10 <sup>8</sup>	8.57	3.7x10 <sup>7</sup>	7.57
10 <sup>-7</sup>	40	44				
Is Log N between 8.17 and 8.70: <b>Yes</b>						
Is Log N <sub>0</sub> between 7.17 and 7.70: <b>Yes</b>						
Control of weighted mean counts: <b>7.77</b>						

**Table 6: Test Log Reduction Values**

Product Concentration	Vc1	Vc2	Na (Ave Vc1&Vc2x10)	Log Na	Log Reduction (No:7.54)
1ppm	>330	>330	>3300	>3.52	<4.05
200ppm	<14	<14	<140	<2.15	>5.42

**Enterococcus hirae ATCC 10541**

**Table 7: N and N<sub>0</sub> values**

Dilution	Vc1	Vc2	Average N(wm)	Log N	N <sub>0</sub> (N/10)	Log N <sub>0</sub>
10 <sup>-6</sup>	298	281	3.1x10 <sup>8</sup>	8.49	3.1x10 <sup>7</sup>	7.49
10 <sup>-7</sup>	36	30				
Is Log N between 8.17 and 8.70: <b>Yes</b>						
Is Log N <sub>0</sub> between 7.17 and 7.70: <b>Yes</b>						
Control of weighted mean counts: <b>8.77</b>						

TECHNICAL SIGNATORY  
 T. NKABAZA

*[Handwritten Signature]*  
 .....  
 Authorised Signature (original blue ink)

All services are rendered in accordance with Bureau Veritas M&L Laboratory Services (Pty) Ltd General Terms and conditions of Business, which has been supplied to you, this certificate cannot be reproduced except in full without the written consent of M and L Laboratory Services.



M&L LABS

## Certificate/Report

RESULTS REPORTED RELATED ONLY TO ITEMS TESTED

Table 8: Test Log Reduction Values

Product Concentration	Vc1	Vc2	Na (Ave Vc1&Vc2x10)	Log Na	Log Reduction (No:7.49)
1ppm	>330	>330	>3300	>3.52	<3.97
200ppm	<14	<14	<140	<2.15	>5.34

### VALIDATIONS AND CONTROLS

Table 9: *Enterococcus hirae* ATCC 10541

Validation suspension			Experimental Conditions Control A			Neutralizer Control B			Method Validation C (Neat product concentration)		
		Ave			Ave			Ave			Ave
Vc1	59	53	Vc1	38	41	Vc1	45	45	Vc1	39	41
Vc2	47		Vc2	44		Vc2	45		Vc2	43	
0.5x53 (Nvo)= 26.5											
Is the Nvo value between 30-160: <b>YES</b>											
Is the Experimental Condition A ≥ 0.5xNvo value: <b>YES</b>											
Is the Neutralizer Condition B ≥ 0.5xNvo value: <b>YES</b>											
Is the Method Validation C ≥ 0.5xNvo value: <b>YES</b>											

Table 10: *Escherichia coli* ATCC 10536

Validation suspension			Experimental Conditions Control A			Neutralizer Control B			Method Validation C (Neat product concentration)		
		Ave			Ave			Ave			Ave
Vc1	51	48	Vc1	36	37.5	Vc1	62	64	Vc1	53	50.5
Vc2	45		Vc2	39		Vc2	66		Vc2	48	
0.5x48 (Nvo)= 24											
Is the Nvo value between 30-160: <b>YES</b>											
Is the Experimental Condition A ≥ 0.5xNvo value: <b>YES</b>											
Is the Neutralizer Condition B ≥ 0.5xNvo value: <b>YES</b>											
Is the Method Validation C ≥ 0.5xNvo value: <b>YES</b>											

TECHNICAL SIGNATORY  
 T. N. CARBONZA

  
 .....  
 Authorised Signature (original blue ink)

All services are rendered in accordance with Bureau Veritas M&L Laboratory Services (Pty) Ltd General Terms and conditions of Business, which has been supplied to you, this certificate cannot be reproduced except in full without the written consent of M and L Laboratory Services.



M&L LABS

## Certificate/Report

RESULTS REPORTED RELATED ONLY TO ITEMS TESTED

Ref. No : 11125027  
 Issued at : Johannesburg  
 Date : 13/07/2020  
 Page : 5 of 8

**Table 11:** *Pseudomonas aeruginosa* ATCC 15442

Validation suspension		Experimental Conditions Control A		Neutralizer Control B		Method Validation C (Neat product concentration)	
	Ave		Ave		Ave		Ave
Vc1	52	Vc1	74	Vc1	62	Vc1	44
Vc2	46	Vc2	66	Vc2	52	Vc2	39
49		70		57		41.5	
0.5x49 (Nvo)= 24.5							
Is the Nvo value between 30-160: <b>YES</b>							
Is the Experimental Condition A ≥ 0.5xNvo value: <b>YES</b>							
Is the Neutralizer Condition B ≥ 0.5xNvo value: <b>YES</b>							
Is the Method Validation C ≥ 0.5xNvo value: <b>YES</b>							

**Table 12:** *Staphylococcus aureus* ATCC 6538

Validation suspension		Experimental Conditions Control A		Neutralizer Control B		Method Validation C (Neat product concentration)	
	Ave		Ave		Ave		Ave
Vc1	70	Vc1	64	Vc1	42	Vc1	72
Vc2	78	Vc2	66	Vc2	48	Vc2	67
74		65		45		69.5	
0.5x 72 (Nvo)= 36							
Is the Nvo value between 30-160: <b>YES</b>							
Is the Experimental Condition A ≥ 0.5xNvo value: <b>YES</b>							
Is the Neutralizer Condition B ≥ 0.5xNvo value: <b>YES</b>							
Is the Method Validation C ≥ 0.5xNvo value: <b>YES</b>							

### Conclusion

The product tested at **200ppm complied** with the criteria indicated under the "Pass Requirements" of SANS 51276 (EN 1276) standard (obligatory conditions) which requires at least a 99.999% kill (5 log reduction).

TECHNICAL SIGNATORY  
 T. N. G. ABIZA

.....

Authorised Signature (original blue ink)

All services are rendered in accordance with Bureau Veritas M&L Laboratory Services (Pty) Ltd General Terms and conditions of Business, which has been supplied to you, this certificate cannot be reproduced except in full without the written consent of M and L Laboratory Services.